

(g) An owner or operator of a sterilization facility seeking to demonstrate compliance with the standards found at § 63.362(c), (d), or (e) with a control device other than an acid-water scrubber or catalytic or thermal oxidation unit shall provide to the Administrator the information requested under § 63.363(f). The owner or operator shall submit: a description of the device; test results collected in accordance with § 63.363(f) verifying the performance of the device for controlling ethylene oxide emissions to the atmosphere to the levels required by the applicable standards; the appropriate operating parameters that will be monitored; and the frequency of measuring and recording to establish continuous compliance with the standards. The monitoring plan submitted identifying the compliance monitoring is subject to the Administrator's approval. The owner or operator of the sterilization facility shall install, calibrate, operate, and maintain the monitor(s) approved by the Administrator based on the information submitted by the owner or operator. The owner or operator shall include in the information submitted to the Administrator proposed performance specifications and quality assurance procedures for their monitors. The Administrator may request further information and shall approve appropriate test methods and procedures.

(h) An owner or operator of a sterilization facility seeking to demonstrate compliance with the standards found at §§ 63.362(d) or (e) with a monitoring device or procedure other than a gas chromatograph shall provide to the Administrator information describing the operation of the monitoring device or procedure and the parameter(s) that would indicate proper operation and maintenance of the device or procedure. The Administrator may request further information and will specify appropriate test methods and procedures.

§ 63.366 Reporting requirements.

(a) The owner or operator of a source subject to the emissions standards in § 63.362 shall fulfill all reporting requirements in §§ 63.10(a), (d), (e), and (f) of subpart A, according to the applica-

bility in Table 1 of § 63.360. These reports will be made to the Administrator at the appropriate address identified in § 63.13 of subpart A of this part.

(1) Reports required by subpart A and this section may be sent by U.S. mail, fax, or by another courier.

(i) Submittals sent by U.S. mail shall be postmarked on or before the specified date.

(ii) Submittals sent by other methods shall be received by the Administrator on or before the specified date.

(2) If acceptable to both the Administrator and the owner or operator of a source, reports may be submitted on electronic media.

(3) Content and submittal dates for excess emissions and monitoring system performance reports. All excess emissions and monitoring system performance reports and all summary reports, if required per § 63.10(e)(3) (vii) and (viii) of subpart A of this part, shall be delivered or postmarked within 30 days following the end of each calendar half or quarter as appropriate (see § 63.10(e)(3) (i) through (iv) for applicability). Written reports of excess emissions or exceedances of process or control system parameters shall include all information required in § 63.10(c) (5) through (13) of subpart A of this part as applicable in Table 1 of § 63.360 and information from any calibration tests in which the monitoring equipment is not in compliance with PS-9 or the method used for temperature calibration. The written report shall also include the name, title, and signature of the responsible official who is certifying the accuracy of the report. When no excess emissions or exceedances have occurred or monitoring equipment has not been inoperative, repaired, or adjusted, such information shall be stated in the report.

(b) *Construction and reconstruction.* The owner or operator of each source using 10 tons shall fulfill all requirements for construction or reconstruction of a source in § 63.5 of subpart A of this part, according to the applicability in Table 1 of § 63.360, and in this paragraph.

(1) *Applicability.* (i) This paragraph and § 63.5 of subpart A of this part implement the preconstruction review requirements of section 112(i)(1) for